

When you need to measure temperature from multiple channels with precision, you want high accuracy and minimal variation from one channel to another.

The TEMPpoint instrument has a simultaneous architecture, with an independent 24-bit Delta-Sigma converter for each channel. (Delta-Sigma converters provide a built-in anti-aliasing filtering that rejects 50 Hz and 60 Hz power line frequency components). When the instrument gets a pulse from the onboard clock, all input signals are sampled at the same time. This architecture virtually eliminates time skew between channels and simplifies both time and frequency-based analysis techniques. In addition, simultaneous A/Ds eliminate several sources of errors found in multiplexed architectures, including settling time and channel-to-channel crosstalk errors. Further, TEMPpoint instruments provide galvanic isolation of 1000 V to protect each channel from noise and high voltages that are inherent in industrial machinery. All of these factors contribute to accurate temperature measurements across channels.



To characterize the channel-to-channel accuracy of the TEMPpoint RTD instrument, we selected a resistor that approximated a standard European PT1000 sensor at room temperature and placed it across one channel. We then measured its resistance (R) using an Agilent Model 3458 precision 28-bit multimeter; the resulting value was 1095.95 ohms. Since the TEMPpoint instrument provides a constant current (I) of 425  $\mu$ A, we could then compute the corresponding voltage using Ohms law ( $V=I*R$ ).

Using the standard Callendar-van Dusen coefficients for a PT1000 RTD type, we then determined the corresponding temperature (24.64° C) for this voltage.

We then placed this resistor across each of the 48 channels and measured data from all channels for 5 seconds, saving the data to disk. The table on page 3 shows the results of the logged data on each channel compared to the expected value of 24.64° C.

As you can see, the lowest minimum value was 24.57° C on channel 23 for a difference of -0.07° C from the expected value of 24.64° C; the highest maximum value was 26.65 on channel 0 for a difference of +0.01° C from the expected value of 24.64° C. The standard deviation on each channel was no greater than 0.01° C. This data validated the excellent channel-to-channel accuracy of the TEMPpoint RTD instrument.

The following page displays a table with Channel-to-Channel Accuracy of TEMPpoint RTD Instruments.

**Table: Channel-to-Channel Accuracy of TEMPoint  
RTD Instruments**

	<b>Ch0</b>	<b>Ch1</b>	<b>Ch2</b>	<b>Ch3</b>	<b>Ch4</b>	<b>Ch5</b>	<b>Ch6</b>	<b>Ch7</b>
Min	24.62	24.61	24.61	24.6	24.61	24.6	24.62	24.59
Max	24.65	24.63	24.62	24.62	24.63	24.62	24.63	24.61
Avg	24.63	24.62	24.62	24.61	24.62	24.61	24.62	24.60
Std Dev	0.01	0.01	0.00	0.01	0.01	0.01	0.00	0.01
Avg Err	.01	.02	.02	.03	.02	.03	.02	.04
	<b>Ch8</b>	<b>Ch9</b>	<b>Ch10</b>	<b>Ch11</b>	<b>Ch12</b>	<b>Ch13</b>	<b>Ch14</b>	<b>Ch15</b>
Min	24.6	24.6	24.6	24.6	24.6	24.6	24.6	24.58
Max	24.62	24.62	24.63	24.63	24.62	24.62	24.62	24.61
Avg	24.62	24.61	24.61	24.61	24.61	24.61	24.61	24.60
Std Dev	0.01	0.01	0.01	0.01	0.01	0.00	0.01	0.01
Avg Err	.02	.03	.03	.03	.03	.03	.03	.04
	<b>Ch16</b>	<b>Ch17</b>	<b>Ch18</b>	<b>Ch19</b>	<b>Ch20</b>	<b>Ch21</b>	<b>Ch22</b>	<b>Ch23</b>
Min	24.59	24.6	24.6	24.59	24.6	24.59	24.6	24.57
Max	24.61	24.62	24.62	24.61	24.62	24.62	24.62	24.59
Avg	24.60	24.61	24.61	24.61	24.61	24.61	24.61	24.59
Std Dev	0.01	0.01	0.01	0.01	0.01	0.01	0.01	0.01
Avg Err	.04	.03	.03	.03	.03	.03	.03	.05
	<b>Ch24</b>	<b>Ch25</b>	<b>Ch26</b>	<b>Ch27</b>	<b>Ch28</b>	<b>Ch29</b>	<b>Ch30</b>	<b>Ch31</b>
Min	24.6	24.61	24.6	24.6	24.6	24.59	24.59	24.59
Max	24.62	24.63	24.62	24.62	24.62	24.62	24.62	24.62
Avg	24.61	24.62	24.61	24.61	24.61	24.60	24.61	24.61
Std Dev	0.01	0.01	0.01	0.01	0.01	0.01	0.01	0.01
Avg Err	.03	.02	.03	.03	.03	.04	.03	.03
	<b>Ch32</b>	<b>Ch33</b>	<b>Ch34</b>	<b>Ch35</b>	<b>Ch36</b>	<b>Ch37</b>	<b>Ch38</b>	<b>Ch39</b>
Min	24.6	24.6	24.6	24.6	24.59	24.6	24.6	24.59
Max	24.62	24.62	24.62	24.62	24.62	24.62	24.62	24.61
Avg	24.61	24.61	24.61	24.61	24.60	24.61	24.61	24.60
Std Dev	0.01	0.01	0.01	0.01	0.01	0.00	0.01	0.01
Avg Err	.03	.03	.03	.03	.04	.03	.03	.04
	<b>Ch40</b>	<b>Ch41</b>	<b>Ch42</b>	<b>Ch43</b>	<b>Ch44</b>	<b>Ch45</b>	<b>Ch46</b>	<b>Ch47</b>
Min	24.61	24.6	24.6	24.59	24.6	24.6	24.59	24.61
Max	24.63	24.62	24.62	24.61	24.62	24.62	24.62	24.63
Avg	24.62	24.61	24.61	24.61	24.61	24.61	24.61	24.62
Std Dev	0.01	0.01	0.01	0.01	0.01	0.01	0.01	0.01
Avg Err	.02	.03	.03	.03	.03	.03	.03	.02

Absolute Min: 24.57 Absolute Max: 24.65 Max Average Err: .05