

Declaration of Conformity

According to ISO/IEC Guide 22 and EN 45014 Part #: 114-0740 Rev 1-99

Manufacturer's Name: IOtech, Inc.
Manufacturer's Address: 25971 Cannon Road
Cleveland, Ohio 44146
USA

Declares that the product:

Product Name: Extender488 and Extender488/F
Description: IEEE 488.2 Bus Extender

Conforms (under conditions on reverse side) to the following standards:

Safety: EN 61010-1 1993
EMC: CISPR22:1987
EN 55022: 1988 class A
IEC 801-2: 1984/prEN50082-1:1991±8kV AD, criterion A
IEC 801-3: 1984/prEN50082-1:1991-3V/m, criterion A
IEC 801-4: 1988/prEN50082-1:1991±0.25kV signal ±0.5 kV line, criterion A

Place: Cleveland, Ohio USA
Date: 1/1999

Signature: 
Full Name: Paul Wittibschlager
Position: Director of Hardware Engineering

European Contact: _____

CE Compliant Operating Conditions

Product Name: Extender488 and Extender488/F
Description: IEEE 488.2 Bus Extender

To maintain safety, emission, and immunity standards of this declaration, the following conditions must be met.

- * IEEE cables must have a braided shield connected circumferentially to their connectors' metal shells.
 - * All cable screw locks must be tightened at both ends of the cable.
 - * The host computer must be properly grounded.
 - * Some inaccuracy is to be expected when I/O leads are exposed to RF fields or transients.
-  The operator must observe all safety cautions and operating conditions specified in the documentation for all hardware used.
-  The host computer, peripheral equipment, power sources, and expansion hardware must be CE compliant.



WARNING. Noted conditions pertain to potential safety hazards. When you see this symbol on the product or in the documentation, carefully read the related information and be alert to the possibility of personal injury.